



525 Series
Σειρά 525
525 Serien

525 Serie
525 Sarja
525 Serie

Série 525
Séria 525
525 系列



en DeVilbiss® 5-Liter Oxygen Concentrator Instruction Guide

WARNING—Read instruction guide before operating this equipment.

ASSEMBLED IN USA



el Έντυπο καθοδήγησης για τον συγκεντρωτή 5 λίτρων Οξυγόνο DeVilbiss®

ΠΡΟΕΙΔΟΠΟΙΗΣΗ — Διαβάστε τον οδηγό λειτουργίας πριν χρησιμοποιήσετε τον εξοπλισμό

ΣΥΝΑΡΜΟΛΟΓΕΙΤΑΙ ΣΤΙΣ ΗΠΑ



da Brugsvejledning til DeVilbiss® 5 liter- iltkoncentrator

ADVARSEL — Læs brugsvejledningen, før udstyret tages i brug.

SAMLET I USA



sv Användarhandledning för DeVilbiss® 5 Liter oxygenkoncentrator

VARNING — Läs bruksanvisningen innan du använder denna utrustning.

MONTERAD I U.S.A.



fi DeVilbiss® 5 litre kannettavan happijärjestelmän käyttöopas

VAROITUS — Lue käyttöohje ennen tämän laitteen käyttämistä.

KOOTTU YHDYSVALLOISSA



no DeVilbiss® 5 Liter Oxygen Concentrator (oksygenkonsentrator) bruksanvisning

ADVARSEL — Les instruksjonshåndboken før bruk av utstyret.


MONTERT I USA



cs Koncentrátor kyslíku DeVilbiss® 5L: Návod k použití

VAROVÁNÍ — Než uvedete toto zařízení do provozu, přečtěte si návod k použití.

SESTAVENO V USA



sk Návod na používanie 5-litrového kyslíkového koncentratora DeVilbiss®

VAROVANIE — Pred použitím tohto zariadenia si prečítajte návod na použitie.

MONTOVANÉ V USA



zh DeVilbiss® 5 升医用制氧机使用指南

警告 — 在操作本设备前，请阅读使用指南。

美国组装



ENGLISH.....	en-2
ΕΛΛΗΝΙΚΑ.....	el-14
DANISH.....	da-27
SWEDISH.....	sv-40
FINNISH.....	fi-52
NORWEGIAN.....	no-64
CZECH.....	cs-76
SLOVAK.....	sk-88
CHINESE (SIMPLIFIED).....	zh-101

TABLE OF CONTENTS

Symbol Definitions	en - 3
Important Safeguards	en - 3
Introduction.....	en - 5
Intended Use	en - 5
Indications For Use	en - 5
Contraindications	en - 5
Essential Performance	en - 5
Service Life	en - 5
Why Your Physician Prescribed Supplemental Oxygen	en - 5
How Your Concentrator Works	en - 5
Important Parts of Your Concentrator.....	en - 6
Setting Up Your Concentrator	en - 7
Operating Your Concentrator	en - 7
DeVilbiss OSD® Operation.....	en - 8
Reserve Oxygen System	en - 8
Caring for Your Concentrator	en - 8
Troubleshooting	en - 9
Overview of Alarms	en - 10
Return and Disposal.....	en - 10
Provider's Notes.....	en - 10
Specifications	en - 11
Electromagnetic Compatibility Information	en - 12
Warranty	en - 13



WARNING

Under certain circumstances, oxygen therapy can be hazardous. Seek medical advice before using an oxygen concentrator.

Physician Information

Physician Name: _____

Telephone: _____

Address: _____

Prescription Information

Name: _____

Oxygen liters per minute

at rest: _____ during activity: _____ other: _____

Oxygen use per day

Hours: _____ Minutes: _____

Comments: _____


DeVilbiss 5-Liter Oxygen Concentrator w/OSD Serial Number: _____

DeVilbiss Equipment Provider Information

Set-Up Person: _____


































This instruction guide was reviewed with me, and I have been instructed on the safe use and care of the DeVilbiss Oxygen Concentrator.

Signature: _____ Date: _____



DeVilbiss 5-Liter Series

SYMBOL DEFINITIONS

	It is mandatory to read and understand the operating instructions prior to use. i This symbol has a blue background on the product label.		Off On		LOT Number		Manufacturer
	Electric Shock Hazard. Cabinet to be removed by authorized personnel only. i This symbol has a yellow background on the product label.		Reset		Catalog Number		European Representative
	Danger - No smoking near patient or device. i This symbol has a red circle and diagonal bar on the product label.		Alternating Current		Serial Number		European CE mark
	Use no Oil, Grease or Lubricants i This symbol has a red circle and diagonal bar on the product label.		Type B applied part		Normal Oxygen		Keep unit dry.
	Do not use near heat or open flames i This symbol has a red circle and diagonal bar on the product label.		Double Insulated		Low Oxygen	 ≤ 5 LPM Maximum recommended flow rate: 5LPM	Maximum recommended flow rate: 5LPM
	General Warning i This symbol is used throughout this manual to indicate hazardous situations to avoid.		Hour Meter		Service Required		
	Important Information i This symbol is used throughout this manual to indicate important information you should know.		Operating Temperature Range +5 to +35°C (+41 to +95°F)		TUV Rheinland C-US approval mark		
	Note and Information Symbol i This symbol is used throughout this manual to indicate notes, useful tips, recommendations and information.		Atmospheric Pressure Range 840 to 1010 hPa (Approximate sea level to 5000 ft)		TUV Rheinland Certified approval mark		Inmetro approval mark
	Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.				Ingress Protection - Protected against finger access to hazardous parts; protected against vertically falling water drops.		
	This device contains electrical and/or electronic equipment that must be recycled per EU Directive 2012/19/EU- Waste Electrical and Electronic Equipment (WEEE)						

IMPORTANT SAFEGUARDS

Read this entire guide before using your DeVilbiss concentrator. Important safeguards are indicated throughout this guide. Pay special attention to all safety information. Imminently and potentially hazardous information is highlighted by these terms:

! DANGER
Indicates an imminently hazardous situation which could result in death or serious injury to the user or operator if not avoided.

! WARNING
Indicates a potentially hazardous situation which could result in death or serious injury to the user or operator if not avoided.

! CAUTION
Indicates a potentially hazardous situation which could result in property damage, injury, or device damage if not avoided.

! IMPORTANT
Indicates important information you should know.

i NOTE
Indicates notes, useful tips, recommendations, and information.

READ ALL INSTRUCTIONS BEFORE USING.



DANGER

- Oxygen causes rapid burning. Do not smoke while your oxygen concentrator is operating, or when you are near a person utilizing oxygen therapy.
- Smoking during oxygen therapy is dangerous and is likely to result in facial burns or death. Do not allow smoking within the same room where the oxygen concentrator or any oxygen carrying accessories are located.
 - If you intend to smoke, you must always turn the oxygen concentrator off, remove the cannula and leave the room where either the cannula or mask or the oxygen concentrator is located. If unable to leave the room, you must wait 10 minutes after you have turned off the oxygen concentrator before smoking.
- Oxygen makes it easier for a fire to start and spread. Do not leave the nasal cannula or mask on bed coverings or chair cushions if the oxygen concentrator is turned on but not in use. The oxygen will make the materials flammable. Turn the oxygen concentrator off when not in use to prevent oxygen enrichment.
- Keep the oxygen concentrator and cannula at least 2 m (6.5 feet) from hot, sparking objects or naked sources of flame.
- Open flames during oxygen therapy are dangerous and are likely to result in fire or death. Do not allow open flames within 2 m (6.5 feet) of the oxygen concentrator or any oxygen carrying accessories.
- DeVilbiss oxygen concentrators are equipped with a fire mitigating outlet fitting that prevents propagation of fire into the unit.



WARNING

- To avoid electric shock, do not plug the concentrator into an AC outlet if the concentrator cabinet is broken. Do not remove the concentrator cabinet. The cabinet should only be removed by a qualified DeVilbiss technician. Do not apply liquid directly to the cabinet or utilize any petroleum-based solvents or cleaning agents.
- Improper use of the power cord and plugs can cause a burn, fire or other electric shock hazards. Do not use the unit if the power cord is damaged.
- Ensure the mains power cord is fully inserted into the concentrator connector (230 volt units) and the power cord plug is completely inserted into a fully functioning AC wall outlet. Failure to do so may cause an electrical safety hazard.
- In order to prevent a fire propagating from the patient through the cannula towards the unit, a means of protection should be located as close to the patient as practicable. Country Standards may vary Please contact your provider for information
- Locate oxygen tubing and power supply cords to prevent tripping hazards and reduce the possibility of entanglement or strangulation.
- Do not lubricate fittings, connections, tubing or other accessories of the oxygen concentrator to avoid the risk of fire and burns.
- Do NOT use lubricants, oils or grease.
- Before attempting any cleaning procedures, turn the unit "Off."
- Use only water-based lotions or salves that are oxygen-compatible before and during oxygen therapy. Never use petroleum or oil-based lotions or salves to avoid the risk of fire and burns.
- Use only spare parts recommended by the manufacturer to ensure proper function and to avoid the risk of fire and burns.
- When using the Transfiller Caddy with a Transfill device, always keep the system on a flat surface. Disassemble the system prior to moving.



WARNING

- If you feel discomfort or are experiencing a medical emergency while undergoing oxygen therapy, seek medical assistance immediately to avoid harm.
- Geriatric, pediatric or any other patient unable to communicate discomfort can require additional monitoring and/or a distributed alarm system to convey the information about the discomfort and/or the medical urgency to the responsible caregiver to avoid harm.
- Use of this device at an altitude above 13,123 feet (4000 meters) or above a temperature of 95°F (35°C) or greater than 93% relative humidity may affect the flow rate and the percentage of oxygen and consequently the quality of the therapy. Refer to specifications for details regarding parameters tested.
- To ensure you receive the therapeutic amount of oxygen delivery according to your medical condition, the Oxygen Concentrator must:
 - be used only after one or more settings have been individually determined or prescribed for you at your specific activity levels.
 - be used with the specific combination of parts and accessories that are in line with the specification of the concentrator manufacturer and that were used while your settings were determined.
- Your delivery settings of the oxygen concentrator should be periodically reassessed for the effectiveness of therapy.
- For your safety, the oxygen concentrator must be used according to the prescription determined by your physician.
- Under certain circumstances, oxygen therapy can be hazardous. Seek medical advice before using an oxygen concentrator.



WARNING

MR Unsafe

- Do not bring the device or accessories into a Magnetic Resonance (MR) environment as it may cause unacceptable risk to the patient or damage to the oxygen concentrator or MR medical devices. The device and accessories have not been evaluated for safety in an MR environment.
- Do not use the device or accessories in an environment with electromagnetic equipment such as CT scanners, Diathermy, RFID and electromagnetic security systems (metal detectors) as it may cause unacceptable risk to the patient or damage to the oxygen concentrator. Some electromagnetic sources may not be apparent, if you notice any unexplained changes in the performance of this device, if it is making unusual or harsh sounds, disconnect the power cord and discontinue use. Contact your home care provider.
- This device is suitable for use in home and healthcare environments except for near active HF SURGICAL EQUIPMENT and the RF shielded room of an ME SYSTEM for magnetic resonance imaging, where the intensity of Electromagnetic DISTURBANCES is high.
- Use of this equipment adjacent to or stacked with other equipment should be avoided because it could result in improper operation. If such use is necessary, this equipment and the other equipment should be observed to verify that they are operating normally.
- Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the oxygen concentrator, including cables specified by the manufacturer. Otherwise, degradation of the performance of this equipment could result.



CAUTION

- It is very important to follow your oxygen prescription. Do not increase or decrease the flow of oxygen – consult your physician.
- The surface temperature of the exhaust vents on the bottom of the unit may exceed 105.8°F (41°C) under certain conditions.
- When device is used under extreme operating conditions, the temperature near the exhaust vents on the bottom of the unit may reach 145°F (63°C). Keep body parts a minimum of 32" (81.2 cm) away from this area.
- Use of harsh chemicals (including alcohol) is not recommended. If bactericidal cleaning is required, a non-alcohol based product should be used to avoid inadvertent damage.



IMPORTANT

- It is recommended that the homecare provider lock the flow control knob to prevent inadvertent adjustment. A flow setting other than prescribed may affect the patient therapy.
- Do not service or clean this device while in use with a Patient.
- Installation of 515LF-607 low output flow meter package will cause the low flow alarm to not work and will prevent the device from meeting the requirements of ISO-80601-2-69:2014 Section 201.13.2.101.
- The Device is classified as IP21 which means it is protected against finger access to hazardous parts and protected against vertically falling water drops.
- Equipment not suitable for use in the presence of a flammable anesthetic mixture with air or with oxygen or nitrous oxide.
- This device contains electrical and/or electronic equipment. Follow local governing ordinances and recycling plans regarding disposal of device components.

SAVE THESE INSTRUCTIONS.

INTRODUCTION

This instruction guide will acquaint you with your DeVilbiss oxygen concentrator. Make sure that you read and understand this guide before operating your unit. Important safeguards are indicated throughout this guide. Pay special attention to all safety information. Contact your DeVilbiss equipment provider should you have any questions.

Intended Use

The DeVilbiss 5 Liter Oxygen Concentrator intended use is to provide supplemental low flow oxygen therapy for patients suffering from COPD, cardiovascular disease, and lung disorders. The oxygen concentrator is used in home type environments, homes, nursing homes, patient care facilities, etc.

Indications For Use

The DeVilbiss Oxygen Concentrator is intended for use as an oxygen concentrator to provide supplemental low flow oxygen therapy in the home, nursing homes, patient care facilities, etc.

Contraindications

The device is not intended to be life supporting or life sustaining.

Essential Performance

Essential Performance of the Oxygen Concentrator is to deliver a continuous flow of oxygen enriched gas. Visual and audible alarms indicate if the device is not meeting specification or a failure has been detected.

Service Life

The expected service life of the 525 is 5 years of operation, when used in accordance with all manufacturer guidance for safe use, maintenance, storage, handling and general operation. Expected service life of the unit, and in particular the sieve beds and compressor, may vary based on the operating environment, storage, handling and the frequency and intensity of use.

Why Your Physician Prescribed Supplemental Oxygen

Today, many people suffer from heart, lung and other respiratory diseases. Many of these people can benefit from supplemental oxygen therapy. Your body requires a steady supply of oxygen to function properly. Your physician prescribed supplemental oxygen for you, because you are not getting enough oxygen from room air alone. Supplemental oxygen will increase the amount of oxygen that your body receives.

Supplemental oxygen is not addictive. Your physician prescribed a specific oxygen flow to improve symptoms such as headaches, drowsiness, confusion, fatigue or increased irritability. If these symptoms persist after you begin your supplemental oxygen program, consult your physician.

The oxygen delivery setting has to be determined for each patient individually with the configuration of the equipment to be used, including accessories.

The proper placement and positioning of the prongs of the nasal cannula in the nose is critical to the amount of oxygen delivered to the respiratory system of the patient.


Your Delivery settings of the oxygen concentrator should be periodically reassessed for the effectiveness of therapy.

How Your DeVilbiss Oxygen Concentrator Works

Oxygen concentrators are the most reliable, efficient and convenient source of supplemental oxygen available today. The oxygen concentrator is electrically operated. The unit separates oxygen from room air which allows high-purity supplemental oxygen to be delivered to you through the oxygen outlet. Although the concentrator filters the oxygen in a room, it will not affect the normal amount of oxygen in your room.

IMPORTANT PARTS OF YOUR CONCENTRATOR


Please take time to familiarize yourself with your DeVilbiss oxygen concentrator before operating.



Front View (Figure A)

- 1. Operating instructions
- 2. Power Switch
| = ON
O = OFF
- 3. Flow meter knob
- 4. Flow meter
- 5. Circuit breaker – resets the unit after electrical overload shutdown
- 6. Oxygen outlet – oxygen is dispersed through this port
- 7. Normal Oxygen (green) light (see page 8)
- 8. Low Oxygen (yellow) light (see page 8)
- 9. Red Service Required (red) light – when illuminated contact your DeVilbiss provider
- 10. Hour meter

Back View (Figure B)


- 11. Handgrip
 - 12. Exhaust
- 

CAUTION
When device is used under extreme operating conditions, the temperature near the exhaust vents on the bottom of the unit may reach 145°F (63°C). Keep body parts a minimum of 32" (81.2 cm) away from this area.
- 13. Power cord and/or IEC power connector.
 - 14. Line cord strap
 - 15. Filter Door with venting and compartment for optional gross particle filter.
 - 16. Auxiliary Oxygen Port (Serial numbers starting with R, N, or B): Your concentrator is equipped with an auxiliary oxygen port that can be used to fill oxygen cylinders with an FDA-cleared cylinder filling device that is designed to use oxygen from a concentrator to fill a cylinder. The port is only for use with FDA-cleared filling devices with compatible oxygen input specifications. Refer to the cylinder filling device instruction guide for the oxygen input/output specifications, connection and operating instructions.

Accessories

- Transfiller CaddyDeVilbiss 525DD-650
- Bubble HumidifierSalter Labs 7600 or equivalent

There are many types of humidifiers, oxygen tubing and cannulas/masks that can be used with this device. Certain humidifiers and accessories may impair the device's performance. A mask or any nasal cannula can be used with continuous flow delivery and may be sized according to your prescription as recommended by your homecare provider who should also give you advice on the proper usage, maintenance and cleaning.



WARNING
When using the Transfiller Caddy with a Transfill device, always keep the system on a flat surface. Disassemble the system prior to moving.

NOTE – The bubble humidifier should be supplied with a permanent fire stop device. If a bubble humidifier needs to be used without a permanent fire stop device, a secondary fire stop device must be used and placed as close to the humidifier as possible. Failing to do so could increase the risk of fire. Country Standards may vary. Please contact your provider for information.

NOTE – A maximum of 50 feet (15 meters) of crush-proof oxygen tubing, plus 7 feet (2.1 meters) of cannula, plus a bubble humidifier is allowed between the concentrator and the patient.

NOTE – The oxygen supply accessory (patient tubing) shall be equipped with a means that, in case of fire, stops the delivery of oxygen to the patient. This means of protection should be located as close to the patient as practicable. Country Standards may vary. Please contact your provider for information.

NOTE – Your healthcare provider should verify the compatibility of the oxygen concentrator and all of the parts used to connect to the patient before use.

SETTING UP YOUR OXYGEN CONCENTRATOR

1. Position your unit near an electrical outlet in the room where you spend most of your time.

NOTE – Do not connect to an electrical outlet controlled by a wall switch.



DANGER

Oxygen causes rapid burning. Do not smoke while your oxygen concentrator is operating, or when you are near a person utilizing oxygen therapy. Keep the oxygen concentrator and cannula at least 2 m (6.5 feet) from hot, sparking objects or naked sources of flame.

2. Position your unit on a flat surface at least 6 inches (16 cm) from walls, draperies or any other objects that might prevent the proper flow of air in and out of your oxygen concentrator. The oxygen concentrator should be located in a well-ventilated area to avoid pollutants or fumes.
NOTE – To move the unit, firmly grasp the handle located on the top of the unit, rolling and/or lifting the unit over pathway obstacles.
3. Before operating your unit, always check to be sure the filter door vents (located on the back of your unit) are clean. Proper cleaning of is discussed in the Caring For Your Concentrator section on page 8.
4. Attach the appropriate oxygen accessories to the oxygen outlet.

Oxygen Tubing Connection:

- a. Thread the supplied oxygen outlet connector onto the oxygen outlet.
- b. Attach the oxygen tubing directly to the connector (Figure 1).

Oxygen Tubing Connection With Humidification:

If your physician has prescribed an oxygen humidifier as part of your therapy, follow these steps (If using a prefill, go to step b.):

- a. Fill the humidifier bottle as per manufacturer's instructions.
- b. Thread the wing nut located on the top of the humidifier bottle to the oxygen outlet so that it is suspended (Figure 2). Make sure it is securely tightened.
- c. Attach the oxygen tubing directly to the humidifier bottle outlet fitting (Figure 3).

NOTE – Your physician has prescribed either a nasal cannula or face mask. In most cases, they are already attached to the oxygen tubing. If not, follow the manufacturer's instructions for attachment.

NOTE – Your healthcare provider should verify the compatibility of the oxygen concentrator and all of the parts used to connect to the patient before use.

5. Remove the power cord completely from the line cord strap. Make sure the power switch is in the "Off" position and insert the plug into the wall outlet. The unit is double insulated to guard against electric shock.



WARNING

Ensure the mains power cord is fully inserted into the concentrator connector (230 volt units) and the power cord plug is completely inserted into a fully functioning AC wall outlet. Failure to do so may cause an electrical safety hazard.

NOTE – (only 115 volt units) The plug on the DeVilbiss oxygen concentrator has one blade wider than the other. To reduce the risk of electric shock, this plug is intended to fit in a wall outlet only one way. Do not attempt to defeat this safety feature.

NOTE – To check your oxygen concentrator and accessories for proper operation; 1. Check the output flow by placing the end of the nasal cannula under the surface of a half-full cup of water and look for the bubbles. 2. Check the system for leaks by bending the nasal prongs over and squeeze tight to stop the flow of oxygen. Look at the flow meter to see that the indicator ball on the flow meter drops to zero. If the indicator ball does not drop to zero, check all connections for possible leaks. Parts to check for leaks are: tubing connections, humidifier bottle and other accessories like firebreaks. Repeat these steps until the flow meter ball drops to zero. Contact your provider or service supplier immediately if you encounter any problems.



WARNING

Improper use of the power cord and plugs can cause a burn, fire or other electric shock hazards. Do not use the unit if the power cord is damaged.

OPERATING YOUR DEVILBISS OXYGEN CONCENTRATOR



DANGER

Oxygen causes rapid burning. Do not smoke while your oxygen concentrator is operating, or when you are near a person utilizing oxygen therapy. Keep the oxygen concentrator and cannula at least 2 m (6.5 feet) from hot, sparking objects or naked sources of flame.

DeVilbiss oxygen concentrators are equipped with a fire mitigating outlet fitting that prevents propagation of fire into the unit.



WARNING

In order to prevent a fire propagating from the patient through the cannula towards the unit, a means of protection should be located as close to the patient as practicable. Please contact your provider for this means of protection.

1. Press the power switch to the "On" position. When the unit is turned "On," all three lights (Service Required, Low Oxygen and Normal Oxygen) on the front panel will illuminate briefly and an audible signal will briefly alarm confirming that the LEDs and audible signal are functioning properly. The unit will then operate in "start up" mode with the Low Oxygen light lit until a normal oxygen level is achieved, at which time the Normal Oxygen light will remain lit. The "start up" may take up to 15 minutes.
NOTE– DeVilbiss recommends for optimal service life that the DeVilbiss Oxygen Concentrator be operated for at least 30 minutes after it is powered on. Shorter periods of operation, operating in extreme temperature/humidity conditions or in the presence of contaminants, and/or handling and storage conditions outside those specified, may affect the long term reliable operation of the product.



**NO
SMOKING**

**DANGER**

Oxygen makes it easier for a fire to start and spread. Do not leave the nasal cannula or mask on bed coverings or chair cushions if the oxygen concentrator is turned on but not in use, the oxygen will make the materials flammable. Turn the oxygen concentrator off when not in use to prevent oxygen enrichment.

NOTE – If the audible signal alarms but the unit is not operating, there is no power to the unit. Refer to the Minor Troubleshooting chart on page 9 and contact your DeVilbiss provider if necessary.

NOTE – If an audible low-frequency vibration sound is detected, the unit is not operating properly. Refer to the Minor Troubleshooting chart on page 9, and contact your DeVilbiss provider if necessary.

2. Check the flow meter to make sure that the flow meter ball is centered on the line next to the prescribed number of your flow rate.

**CAUTION**

It is very important to follow your oxygen prescription. Do not increase or decrease the flow of oxygen – consult your physician.

NOTE – Your DeVilbiss provider may have preset the flow meter so that it cannot be adjusted.

NOTE – If the flow meter knob is turned clockwise, the flow decreases (and eventually will shut off the oxygen flow). If the knob is turned counterclockwise, the flow increases.

NOTE – For prescriptions of 5 LPM, be sure the ball is centered on the 5 liter line. The ball should not touch the red line. Setting the flow higher than 5 may cause the oxygen purity level to drop.

NOTE – The low-flow alarm may activate if the flow meter ball is set at or below 0.2 lpm. The unit will continue to run; however, the Service Required light will come on accompanied by an audible alarm. Adjust the flow meter to your prescribed flow.

3. Your DeVilbiss concentrator is now ready for use, properly position the cannula with the nasal prongs facing upward, insert the prongs into nose. Wrap the cannula tubing over the ears and position in front of body (Figure 4). Allow 15 minutes for the oxygen concentrator to reach stated performance.

**DeVilbiss OSD® Operation**

The OSD (Oxygen Sensing Device) is a device within your concentrator that monitors the oxygen produced by your unit.

The OSD lights on the top panel are defined as follows:

- Green Normal Oxygen light—acceptable oxygen level.
- Yellow Low Oxygen light— below an acceptable oxygen level.

If the oxygen purity falls below the acceptable level: The green Normal Oxygen light will shut off, the yellow Low Oxygen light will illuminate, and an intermittent audible signal will sound.

Refer to the Minor Troubleshooting section in this guide on page 9, and switch to your reserve oxygen system. Do not attempt any other maintenance. Contact your DeVilbiss provider immediately.

RESERVE OXYGEN SYSTEM

As a precaution, your DeVilbiss provider may supply you with a reserve oxygen system. If your unit loses electrical power or fails to operate correctly, the Patient Alert System will sound to signal you to switch to your reserve oxygen system (if provided) and contact your DeVilbiss provider.

CARING FOR YOUR DEVILBISS OXYGEN CONCENTRATOR

DeVilbiss recommends using only original DeVilbiss parts and filters in order to guarantee reliable operation of the product.

**WARNING**

Do NOT use lubricants, oils or grease.
Before attempting any cleaning procedures, turn the unit “Off.”

Cannula/Mask, Tubing and Humidifier Bottle

Clean and replace the cannula/mask, tubing and humidifier bottle according to the manufacturer's instructions.

Filter Door with Vents

Inspect the vents periodically, and wipe with a dry cloth as needed to remove dust.

Exterior Cabinet

Clean the concentrator exterior cabinet weekly by using a damp cloth and wiping dry; the vents can also be wiped with a damp cloth.

Cleaning

	Recommended cleaning interval	Number of cleaning cycles *	Compatible cleaning method
Outer Cabinet	7 days	260	Water, use only a damp cloth
Filter Door Vents	7 days	260	Wipe with dry cloth, or a cloth dampened with water to remove dust.
Oxygen Outlet Connector	7 days	104	Mild dish soap (2 tbsp) and warm water (2 cups)

* number of cleaning cycles determined by recommended cleaning interval and expected service life

**WARNING**

To avoid electric shock, do not plug the concentrator into an AC outlet if the concentrator cabinet is broken. Do not remove the concentrator cabinet. The cabinet should only be removed by a qualified DeVilbiss technician. Do not apply liquid directly to the cabinet or utilize any petroleum-based solvents or cleaning agents.

**CAUTION**

Use of harsh chemicals (including alcohol) is not recommended. If bactericidal cleaning is required, a non-alcohol based product should be used to avoid inadvertent damage.

TROUBLESHOOTING

The following troubleshooting chart will help you analyze and correct minor oxygen concentrator malfunctions. If the suggested procedures do not help, switch to your reserve oxygen system and call your DeVilbiss homecare provider. Do not attempt any other maintenance.

**WARNING**

To avoid electric shock, do not plug the concentrator into an AC outlet if the concentrator cabinet is broken. Do not remove the concentrator cabinet. The cabinet should only be removed by a qualified DeVilbiss technician.

Minor Troubleshooting Chart

SYMPTOM	POSSIBLE CAUSE	REMEDY
A. Unit does not operate. All lights are off when the power switch is "On." Audible alert is pulsing.	1. Power cord not properly inserted into wall outlet.	1. Check power cord connection at the wall outlet. On 230 volt units, also check the mains connection on the back of the unit.
	2. No power at wall outlet.	2. Check your home circuit breaker and reset if necessary. Use a different wall outlet if the situation occurs again.
	3. Oxygen concentrator circuit breaker activated.	3. Press the concentrator circuit breaker reset button located below the power switch. Use a different wall outlet if the situation occurs again. If the above remedies do not work, contact your DeVilbiss provider.
B. Unit operates. Red Service Required light is illuminated. Audible alert may be sounding.	1. Filter door vents are blocked.	1. Check filter door vents and ensure that the openings are not blocked.
	2. Exhaust is blocked.	2. Check the exhaust area and make sure there is nothing restricting the unit exhaust.
	3. Blocked or defective cannula, face mask, or oxygen tubing.	3. Detach cannula or face mask. If proper flow is restored, clean or replace if necessary. Disconnect the oxygen tubing at the oxygen outlet. If proper flow is restored, check oxygen tubing for obstructions or kinks. Replace if necessary.
	4. Blocked or defective humidifier bottle.	4. Detach the humidifier from the oxygen outlet. If proper flow is obtained, clean or replace humidifier.
	5. Flow meter set too low.	5. Set flow meter to prescribed flow rate. If the above remedies do not work, contact your DeVilbiss provider.
C. Unit operates. Audible low-frequency vibration sound is detected.	1. Electronic Assembly Malfunction.	1. Turn your unit "Off." Switch to your reserve oxygen system and contact your DeVilbiss provider immediately.
D. Yellow Low Oxygen light is on.	1. Unit in "start up" mode.	1. Allow unit up to 15 minutes to complete start up period.
E. The yellow Low Oxygen light is on and the intermittent audible signal is sounding.	1. Flow meter is not properly set.	1. Ensure the flow meter is properly set to the prescribed number. (The maximum flow meter setting is 3 LPM when an oxygen bottle is being filled with oxygen from the auxiliary port.)
	2. Filter door vents are blocked.	2. Check filter door vents and ensure that the openings are not blocked.
	3. Exhaust is blocked.	3. Check the exhaust area and make sure there is nothing restricting the unit exhaust. If the above remedies do not work, contact your DeVilbiss provider.
F. Red Service Required light is on and an intermittent audible signal is sounding.	1. Flow meter is not properly set.	1. Ensure the flow meter is properly set to the prescribed number. (The maximum flow meter setting is 3 LPM when an oxygen bottle is being filled with oxygen from the auxiliary port.)
	2. Filter door vents are blocked.	2. Check filter door vents and ensure that the openings are not blocked.
	3. Exhaust is blocked.	3. Check the exhaust area and make sure there is nothing restricting the unit exhaust. If the above remedies do not work, contact your DeVilbiss provider.
	4. Electronic Assembly Malfunction.	4. Turn your unit "Off." Switch to your reserve oxygen system and contact your DeVilbiss provider immediately.
G. If any other problems occur with your oxygen concentrator.		1. Turn your unit "Off." Switch to your reserve oxygen system and contact your DeVilbiss provider immediately.